









Kuwait Anesthesia & Critical Care Society

Compatibility Table 2024



SAD/ETT	<i>Endotracheal tube</i> 						
	Sizes	6.0	6.5	7	7.5	8	8.5
<i>I-GEL</i> 	3	yes	yes	no	no	no	NA
	4	yes	yes	yes	no	no	NA
	5	yes	yes	yes	yes	yes	NA
<i>Ambu</i> 	3	yes	yes	no	no	no	NA
	4	yes	yes	yes	yes	no	NA
	5	yes	yes	yes	yes	yes	NA
<i>Classic</i> 	3	yes	no	no	no	no	NA
	4	yes	no	no	no	no	NA
	5	yes	yes	yes	no	no	NA
<i>Air-Q</i> 	2.5	yes	yes	no	no	no	no
	3.5	yes	yes	yes	yes	no	no
	4.5	yes	yes	yes	yes	yes	yes
<i>Supreme</i> 	All sizes	Non compatible with ETT					

SAD: Supraglottic Airway Device

ETT: Endotracheal Tube

Kindly forward feedback to: KACCS@kma.org.kw

